
Midmark M11 - v1.0.0
Total Cycles: 3366

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2023 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 196.5 F 1.7
4:00 206.5 F 6.7
6:00 222.1 F 11.6
8:00 227.7 F 7.1
10:00 239.6 F 12.3
12:00 238.9 F 10.2
14:00 236.6 F 8.6
16:00 250.7 F 15.6
18:00 262.3 F 22.2
20:00 272.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.3 F 28.9
2:00 272.2 F 28.6
3:00 271.5 F 28.3
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 213.1 F 0.0
10:00 217.4 F 0.0
15:00 215.9 F 0.0
20:00 220.0 F 0.0
25:00 215.9 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

06/09/2023 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____