
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001448
12:56 11/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:17
STERILIZING 4:32
136.4 C 224kPa 4:32
136.3 C 224kPa 5:02
136.4 C 223kPa 5:32
135.9 C 222kPa 6:02
136.4 C 224kPa 6:32
136.3 C 223kPa 7:02
136.4 C 225kPa 7:32
136.3 C 224kPa 8:02
136.3 C 224kPa 8:02
VENTING 8:02
AIR DRYING 8:51
CYCLE COMPLETE 8:57

Digital Signature #
BF2DE502146B86F8
