
Midmark M11 - v1.0.5
Total Cycles: 349

Name: RITTER2
Sterilizer ID: V2313228
09/09/2021 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.3 F 0.0
2:00 190.0 F 3.7
4:00 205.3 F 9.6
6:00 220.7 F 6.7
8:00 235.8 F 12.0
10:00 237.7 F 10.7
12:00 238.2 F 10.1
14:00 252.7 F 17.4
16:00 265.0 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.2 F 29.2
2:00 272.7 F 29.6
3:00 272.8 F 30.0
4:00 271.9 F 29.1

Min 271.8 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 222.0 F 0.0
10:00 221.2 F 0.0
Dry Cycle Aborted

FILLING: 4:24
HEATING: 17:40
STERILIZING: 4:00
VENTING: 3:07
DRYING: 10:05
TOTAL CYCLE: 00:39:17

09/09/2021 03:09 PM
ACCEPTED: X
REJECTED:
Operator: U02