

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1506

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2022 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
2:00 189.9 F 3.1  
4:00 195.6 F 8.5  
6:00 212.9 F 8.5  
8:00 229.7 F 9.9  
10:00 233.1 F 8.5  
12:00 243.0 F 12.3  
14:00 245.1 F 13.3  
16:00 258.0 F 20.2  
18:00 269.4 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.8 F 29.7  
2:00 271.8 F 28.9  
3:00 272.7 F 29.8  
4:00 272.4 F 29.3  
  
Min 271.6 F 28.8  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 213.2 F 0.0  
10:00 224.6 F 0.0  
15:00 227.1 F 0.0  
20:00 227.5 F 0.0  
25:00 226.3 F 0.0  
30:00 186.6 F 0.0  
  
Drying Complete

FILLING: 2:39  
HEATING: 19:02  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:51  
CYCLE COMPLETE

09/20/2022 10:02 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_