
Midmark M11 - v1.0.5
Total Cycles: 1532

Name: RITTER2
Sterilizer ID: V2313228
09/26/2022 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.2 F 0.0
2:00 127.3 F 0.0
4:00 141.2 F 1.7
6:00 166.8 F 5.4
8:00 193.1 F 10.2
10:00 219.0 F 6.2
12:00 225.4 F 9.3
14:00 234.5 F 10.3
16:00 238.3 F 11.7
18:00 237.1 F 10.2
20:00 250.7 F 16.7
22:00 262.0 F 23.2
24:00 272.8 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.2 F 30.4
2:00 272.3 F 30.0
3:00 271.5 F 29.4
4:00 271.4 F 29.9

Min 271.1 F 29.3
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 201.6 F 0.0
10:00 228.4 F 0.0
15:00 232.0 F 0.0
20:00 240.3 F 0.0
25:00 228.2 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

09/26/2022 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____