
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001864
14:58 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:10
STERILIZING 4:22
136.4 C 227kPa 4:22
136.2 C 224kPa 4:52
136.4 C 226kPa 5:22
136.4 C 224kPa 5:52
136.3 C 225kPa 6:22
136.3 C 225kPa 6:52
136.3 C 225kPa 7:22
136.2 C 224kPa 7:52
136.2 C 224kPa 7:52
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 8:50

Digital Signature #
B4EB621A8614F3EA
