
Midmark M11 - v1.0.5
Total Cycles: 463

Name: RITTER2
Sterilizer ID: V2313228
10/20/2021 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.7 F 0.0
2:00 105.3 F 0.0
4:00 110.2 F 0.1
6:00 138.0 F 2.0
8:00 169.8 F 5.9
10:00 191.0 F 10.5
12:00 213.6 F 5.6
14:00 223.6 F 10.0
16:00 229.1 F 8.8
18:00 238.8 F 12.4
20:00 238.3 F 10.7
22:00 250.6 F 16.7
24:00 261.5 F 23.0
26:00 272.0 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.3 F 30.3
2:00 271.6 F 29.7
3:00 271.1 F 29.7
4:00 272.2 F 30.3

Min 271.0 F 29.3
Max 272.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 218.1 F 0.0
10:00 227.7 F 0.0
15:00 226.8 F 0.0
20:00 225.6 F 0.0
25:00 226.3 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:36
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:04:28
CYCLE COMPLETE

10/20/2021 09:47 AM
ACCEPTED: X
REJECTED:
Operator: U01