
Midmark M11 - v1.0.5
Total Cycles: 1563

Name: RITTER2
Sterilizer ID: V2313228
10/04/2022 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 184.1 F 0.1
4:00 201.1 F 2.9
6:00 220.2 F 8.0
8:00 236.0 F 13.5
10:00 236.8 F 10.7
12:00 237.1 F 9.0
14:00 239.4 F 9.7
16:00 250.0 F 15.5
18:00 262.4 F 22.7
20:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.3
2:00 271.5 F 28.8
3:00 273.0 F 29.5
4:00 271.9 F 28.8

Min 271.5 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 225.9 F 0.0
10:00 235.8 F 0.0
15:00 229.7 F 0.0
20:00 228.0 F 0.0
25:00 227.6 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

10/04/2022 03:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____