
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000196
7:28 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:49
STERILIZING 5:21
136.0 C 228kPa 5:21
136.4 C 232kPa 5:51
136.3 C 230kPa 6:21
136.3 C 232kPa 6:51
136.3 C 231kPa 7:21
136.4 C 232kPa 7:51
136.4 C 230kPa 8:21
136.4 C 233kPa 8:51
136.4 C 233kPa 8:51
VENTING 8:51
AIR DRYING 9:52
CYCLE COMPLETE 11:07

Digital Signature #
6EEAEF96D78411B8
