
Midmark M11 - v1.0.0
Total Cycles: 2146

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2022 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 181.6 F 2.2
4:00 195.7 F 6.9
6:00 214.7 F 12.3
8:00 224.4 F 6.9
10:00 236.8 F 12.0
12:00 236.1 F 9.3
14:00 238.4 F 9.3
16:00 248.5 F 14.5
18:00 260.8 F 21.2
20:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.8
2:00 271.4 F 28.1
3:00 272.7 F 28.9
4:00 271.7 F 28.1

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 215.6 F 0.0
10:00 212.7 F 0.0
15:00 213.0 F 0.0
20:00 208.5 F 0.0
25:00 205.7 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

07/18/2022 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____