
Midmark M11 - v1.0.0
Total Cycles: 1539

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2022 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.1 F 0.0
2:00 150.8 F 2.6
4:00 168.0 F 6.6
6:00 192.6 F 11.2
8:00 212.8 F 4.6
10:00 227.0 F 9.7
12:00 233.3 F 8.7
14:00 240.6 F 12.1
16:00 239.6 F 10.5
18:00 253.0 F 17.2
20:00 264.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.1
2:00 272.4 F 29.5
3:00 271.8 F 28.7
4:00 272.6 F 29.5

Min 271.5 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 217.1 F 0.0
10:00 229.9 F 0.0
15:00 232.9 F 0.0
20:00 231.5 F 0.0
25:00 226.3 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

01/19/2022 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____