
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005124
15:14 14/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 2:54
136.4 C 223kPa 2:54
136.0 C 222kPa 3:24
136.2 C 223kPa 3:54
136.4 C 224kPa 4:24
136.3 C 225kPa 4:54
136.4 C 223kPa 5:24
136.4 C 225kPa 5:54
136.2 C 221kPa 6:24
136.2 C 221kPa 6:24
VENTING 6:24
AIR DRYING 7:14
CYCLE COMPLETE 11:45

Digital Signature #
A11B76A4B7885C15
