

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2208

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2022 01:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
2:00 192.4 F 4.4  
4:00 200.2 F 9.4  
6:00 221.1 F 6.9  
8:00 230.8 F 10.7  
10:00 231.7 F 7.5  
12:00 242.6 F 11.8  
14:00 244.7 F 12.7  
16:00 257.2 F 19.4  
18:00 268.5 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.3 F 29.1  
2:00 271.2 F 28.4  
3:00 272.3 F 29.3  
4:00 271.9 F 28.8  
  
Min 271.1 F 28.4  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 229.8 F 0.0  
10:00 240.5 F 0.0  
15:00 236.0 F 0.0  
20:00 228.0 F 0.0  
25:00 226.8 F 0.0  
30:00 232.2 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 19:13  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:12  
CYCLE COMPLETE

08/08/2022 02:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_