
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006087
15:15 09/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:48
STERILIZING 3:47
136.4 C 222kPa 3:47
136.1 C 222kPa 4:17
136.1 C 224kPa 4:47
136.2 C 223kPa 5:17
136.3 C 221kPa 5:47
136.4 C 224kPa 6:17
136.0 C 221kPa 6:47
136.4 C 221kPa 7:17
136.4 C 221kPa 7:17
VENTING 7:18
AIR DRYING 8:10
CYCLE COMPLETE 38:10

Digital Signature #
8C608FBC0239D39E
