
Midmark M11 - v1.0.5
Total Cycles: 1100

Name: RITTER2
Sterilizer ID: V2313228
05/12/2022 06:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 75.1 F 0.0
4:00 83.0 F 0.0
6:00 105.2 F 0.0
8:00 152.7 F 0.2
10:00 177.6 F 1.3
12:00 195.5 F 4.7
14:00 209.6 F 8.2
16:00 221.1 F 11.8
18:00 224.2 F 6.3
20:00 235.6 F 10.8
22:00 236.2 F 8.0
24:00 245.0 F 13.6
26:00 243.8 F 12.3
28:00 255.3 F 18.3
30:00 265.5 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.4 F 28.7
3:00 272.7 F 29.4
4:00 271.6 F 28.6

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 197.5 F 0.0
10:00 207.0 F 0.0
15:00 203.2 F 0.0
20:00 200.1 F 0.0
25:00 204.3 F 0.0
30:00 203.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

05/12/2022 08:15 AM
ACCEPTED: X
REJECTED:
Operator: U02