
Midmark M11 - v1.0.5
Total Cycles: 1274

Name: RITTER2
Sterilizer ID: V2313228
07/08/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.6 F 0.0
2:00 146.3 F 2.6
4:00 171.5 F 6.6
6:00 196.1 F 11.3
8:00 211.3 F 4.5
10:00 224.3 F 9.4
12:00 232.1 F 10.3
14:00 239.9 F 11.7
16:00 238.6 F 10.2
18:00 251.7 F 16.6
20:00 262.9 F 23.1
22:00 272.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 271.9 F 29.3
2:00 271.3 F 29.2
3:00 272.0 F 29.2
4:00 271.2 F 29.1

Min 271.0 F 28.7
Max 272.1 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 224.4 F 0.0
10:00 221.3 F 0.0
15:00 227.8 F 0.0
20:00 236.2 F 0.0
25:00 232.6 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

07/08/2022 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____