
Midmark M11 - v1.0.0
Total Cycles: 2145

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.4 F 0.0
2:00 136.7 F 0.2
4:00 158.7 F 3.5
6:00 190.7 F 7.5
8:00 205.9 F 12.6
10:00 219.1 F 6.1
12:00 232.6 F 10.6
14:00 232.6 F 7.5
16:00 243.3 F 13.1
18:00 241.9 F 11.6
20:00 255.0 F 18.1
22:00 266.0 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.7 F 29.4
2:00 271.8 F 28.7
3:00 271.7 F 29.0
4:00 272.7 F 29.4

Min 271.3 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 236.4 F 0.0
10:00 229.1 F 0.0
15:00 229.1 F 0.0
20:00 219.1 F 0.0
25:00 216.5 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:43
CYCLE COMPLETE

07/18/2022 02:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____