
Midmark M11 - v1.0.0
Total Cycles: 4204

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2024 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 172.6 F 2.7
4:00 182.4 F 6.5
6:00 217.1 F 11.7
8:00 224.8 F 5.8
10:00 232.8 F 10.9
12:00 232.5 F 7.5
14:00 246.2 F 13.6
16:00 244.2 F 12.1
18:00 256.4 F 18.5
20:00 267.2 F 25.1
22:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.0 F 28.4
2:00 272.5 F 28.7
3:00 272.4 F 29.0
4:00 271.5 F 28.3

Min 271.4 F 28.3
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 192.6 F 0.0
10:00 193.6 F 0.0
15:00 193.3 F 0.0
20:00 191.1 F 0.0
25:00 189.8 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

01/15/2024 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____