

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000188  
14:23 06/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:13  
STERILIZING 4:37  
136.0 C 220kPa 4:37  
136.0 C 222kPa 5:07  
136.4 C 222kPa 5:37  
136.2 C 222kPa 6:07  
136.3 C 224kPa 6:37  
136.4 C 225kPa 7:07  
136.2 C 220kPa 7:37  
136.1 C 221kPa 8:07  
136.1 C 221kPa 8:07  
VENTING 8:07  
AIR DRYING 9:10  
CYCLE COMPLETE 9:21  
-----

Digital Signature #  
5CF3FCD70430752E  
-----