
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001669
9:29 01/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:14
STERILIZING 4:21
136.0 C 220kPa 4:21
136.2 C 222kPa 4:51
136.2 C 222kPa 5:21
136.2 C 219kPa 5:51
136.2 C 222kPa 6:21
136.4 C 223kPa 6:51
136.4 C 223kPa 7:21
136.2 C 221kPa 7:51
136.2 C 221kPa 7:51
VENTING 7:51
AIR DRYING 8:45
CYCLE COMPLETE 8:51

Digital Signature #
3E670CF4ECD133A0
