
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002213
16:09 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:33
STERILIZING 3:39
136.0 C 220kPa 3:39
136.2 C 224kPa 4:09
136.2 C 224kPa 4:39
136.2 C 224kPa 5:09
136.3 C 225kPa 5:39
136.3 C 224kPa 6:09
136.4 C 226kPa 6:39
136.2 C 224kPa 7:09
136.2 C 224kPa 7:09
VENTING 7:09
AIR DRYING 8:07
CYCLE COMPLETE 18:49

Digital Signature #
86F307466E2F9AEA
