
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001214
12:44 17/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 4:00
STERILIZING 5:09
136.1 C 221kPa 5:09
136.4 C 224kPa 5:39
136.4 C 224kPa 6:09
136.4 C 224kPa 6:39
136.4 C 225kPa 7:09
136.1 C 222kPa 7:39
136.2 C 222kPa 8:09
136.4 C 225kPa 8:39
136.4 C 225kPa 8:39
VENTING 8:40
AIR DRYING 9:34
CYCLE COMPLETE 24:34

Digital Signature #
ADA8B90562329FBD
