

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002781  
9:43 06/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:06  
STERILIZING 4:22  
136.0 C 220kPa 4:22  
136.4 C 223kPa 4:53  
136.3 C 222kPa 5:23  
136.4 C 221kPa 5:53  
136.2 C 220kPa 6:23  
136.3 C 222kPa 6:53  
136.3 C 223kPa 7:23  
136.2 C 222kPa 7:53  
136.2 C 222kPa 7:53  
VENTING 7:53  
AIR DRYING 8:47  
CYCLE COMPLETE 8:53

-----  
Digital Signature #  
06D98F06D61224A3  
-----