

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1208

Name: RITTER2  
Sterilizer ID: V2313228  
06/21/2022 01:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.4 F 0.0  
2:00 153.7 F 2.7  
4:00 174.7 F 6.5  
6:00 197.2 F 11.5  
8:00 216.6 F 5.2  
10:00 231.6 F 10.1  
12:00 233.3 F 8.5  
14:00 243.7 F 13.0  
16:00 242.7 F 11.9  
18:00 255.8 F 18.7  
20:00 267.1 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.6 F 29.3  
2:00 271.9 F 29.1  
3:00 272.6 F 29.3  
4:00 271.8 F 28.8

Min 271.7 F 28.7  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
5:00 218.5 F 0.0  
10:00 222.9 F 0.0  
15:00 207.1 F 0.0  
20:00 205.0 F 0.0  
25:00 207.5 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:24  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:00:08  
CYCLE COMPLETE

06/21/2022 02:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_