
Midmark M11 - v1.0.5
Total Cycles: 2046

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 164.6 F 0.1
4:00 177.5 F 2.4
6:00 200.0 F 6.7
8:00 217.2 F 11.4
10:00 226.3 F 7.1
12:00 238.5 F 12.6
14:00 238.2 F 10.5
16:00 236.3 F 9.1
18:00 250.6 F 15.9
20:00 262.1 F 22.6
22:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.3
2:00 271.5 F 28.9
3:00 272.6 F 29.4
4:00 271.7 F 28.7

Min 271.4 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 227.9 F 0.0
10:00 240.8 F 0.0
15:00 229.7 F 0.0
20:00 229.8 F 0.0
25:00 227.4 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

02/14/2023 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____