
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.3uS / 1.4ppm
CYCLE NUMBER 002354
8:28 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:21
STERILIZING 3:21
136.4 C 221kPa 3:21
136.3 C 223kPa 3:51
136.0 C 217kPa 4:21
135.9 C 218kPa 4:51
136.1 C 221kPa 5:21
136.3 C 222kPa 5:51
136.4 C 223kPa 6:21
136.2 C 219kPa 6:51
136.2 C 219kPa 6:51
VENTING 6:51
AIR DRYING 7:37
CYCLE COMPLETE 9:41

Digital Signature #
3ABCE9D280E84728
