
Midmark M11 - v1.0.0
Total Cycles: 2813

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2023 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.3 F 0.0
2:00 166.0 F 1.1
4:00 182.6 F 4.9
6:00 206.7 F 10.1
8:00 225.0 F 5.6
10:00 234.7 F 10.1
12:00 236.3 F 8.0
14:00 244.5 F 13.1
16:00 243.4 F 11.9
18:00 256.5 F 18.7
20:00 268.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.3 F 28.8
2:00 271.5 F 28.6
3:00 272.2 F 28.7
4:00 271.4 F 28.4

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 222.1 F 0.0
10:00 228.0 F 0.0
15:00 238.8 F 0.0
20:00 235.9 F 0.0
25:00 228.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

01/24/2023 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____