
Midmark M11 - v1.0.0
Total Cycles: 3798

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2023 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.9 F 0.0
2:00 178.8 F 1.5
4:00 180.7 F 4.9
6:00 210.1 F 9.8
8:00 218.8 F 5.9
10:00 229.0 F 10.1
12:00 232.8 F 7.8
14:00 243.7 F 12.9
16:00 242.5 F 11.5
18:00 255.1 F 17.9
20:00 266.0 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.0 F 28.6
2:00 272.0 F 28.9
3:00 271.6 F 28.3
4:00 272.3 F 28.9

Min 271.3 F 28.2
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 231.6 F 0.0
10:00 216.7 F 0.0
15:00 220.6 F 0.0
20:00 221.7 F 0.0
25:00 230.4 F 0.0
30:00 232.4 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

09/20/2023 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____