
Midmark M11 - v1.0.0
Total Cycles: 3974

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.9 F 0.0
2:00 157.1 F 2.9
4:00 177.6 F 7.1
6:00 200.8 F 12.2
8:00 223.4 F 5.8
10:00 233.4 F 10.8
12:00 231.2 F 6.7
14:00 245.3 F 13.1
16:00 242.7 F 11.5
18:00 255.7 F 18.1
20:00 267.3 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.9 F 28.3
2:00 272.3 F 28.5
3:00 272.4 F 28.8
4:00 272.6 F 29.0

Min 271.5 F 28.3
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 197.2 F 0.0
10:00 250.7 F 0.0
15:00 215.2 F 0.0
20:00 207.3 F 0.0
25:00 214.2 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

11/02/2023 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____