

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5050

Name: RITTER 1  
Sterilizer ID: V2178198  
11/04/2024 12:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.7 F 0.0  
2:00 160.2 F 0.5  
4:00 178.5 F 2.3  
6:00 197.6 F 6.0  
8:00 216.9 F 10.0  
10:00 230.6 F 7.9  
12:00 233.3 F 8.6  
14:00 242.8 F 13.3  
16:00 240.9 F 10.8  
18:00 237.0 F 8.8  
20:00 250.7 F 15.3  
22:00 261.8 F 21.5  
24:00 271.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 27.9  
1:00 272.4 F 28.7  
2:00 272.3 F 28.4  
3:00 272.2 F 28.8  
4:00 272.5 F 28.6

Min 271.0 F 27.8  
Max 273.0 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
5:00 239.5 F 0.0  
10:00 205.8 F 0.0  
15:00 198.3 F 0.0  
20:00 202.6 F 0.0  
25:00 202.7 F 0.0  
30:00 203.8 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:38  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:31  
CYCLE COMPLETE

11/04/2024 01:24 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_