
Midmark M11 - v1.0.5
Total Cycles: 1315

Name: RITTER2
Sterilizer ID: V2313228
07/19/2022 06:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 87.1 F 0.0
4:00 113.0 F 0.0
6:00 127.3 F 0.0
8:00 165.2 F 0.3
10:00 207.0 F 0.9
12:00 215.3 F 3.3
14:00 220.4 F 6.4
16:00 228.9 F 9.7
18:00 232.3 F 12.4
20:00 233.6 F 8.9
22:00 242.8 F 13.5
24:00 241.2 F 11.5
26:00 239.4 F 10.4
28:00 252.3 F 16.7
30:00 263.3 F 23.2
32:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.3 F 29.4
2:00 271.9 F 28.7
3:00 272.4 F 29.5
4:00 272.1 F 28.9

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 227.8 F 0.0
10:00 185.0 F 0.0
15:00 175.2 F 0.0
20:00 168.4 F 0.0
25:00 170.3 F 0.0
30:00 162.9 F 0.0

Drying Complete

FILLING: 0:60
HEATING: 32:07
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:10:11
CYCLE COMPLETE

07/19/2022 08:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____