
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002591
8:08 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:15
STERILIZING 4:19
136.0 C 220kPa 4:19
136.4 C 218kPa 4:49
136.2 C 219kPa 5:19
136.4 C 220kPa 5:49
136.2 C 217kPa 6:19
136.4 C 220kPa 6:49
136.4 C 220kPa 7:19
136.0 C 217kPa 7:49
136.0 C 217kPa 7:49
VENTING 7:49
AIR DRYING 8:40
CYCLE COMPLETE 9:59

Digital Signature #
CDF6F014A9136A6D
