
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000572
15:44 01/10/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 5:00
STERILIZING 5:53
136.4 C 224kPa 5:53
136.3 C 224kPa 6:23
136.3 C 222kPa 6:53
136.4 C 223kPa 7:23
136.3 C 225kPa 7:53
136.3 C 222kPa 8:23
136.4 C 225kPa 8:53
136.3 C 223kPa 9:23
136.5 C 225kPa 9:53
136.3 C 222kPa 10:23
136.3 C 223kPa 10:53
136.4 C 223kPa 11:23
136.3 C 220kPa 11:53
136.3 C 224kPa 12:23
136.4 C 225kPa 12:53
136.3 C 224kPa 13:23
136.4 C 224kPa 13:53
136.1 C 220kPa 14:23
136.3 C 224kPa 14:53
136.1 C 220kPa 15:23
136.3 C 220kPa 15:53
136.3 C 220kPa 15:53
VENTING 15:53
AIR DRYING 16:46
CYCLE COMPLETE 16:57

Digital Signature #

FCDBC28B97E8B340
