
Midmark M11 - v1.0.0
Total Cycles: 2533

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2022 05:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.1 F 0.0
2:00 188.3 F 4.3
4:00 193.8 F 9.8
6:00 212.9 F 4.7
8:00 229.9 F 9.4
10:00 234.8 F 8.3
12:00 245.5 F 13.2
14:00 245.1 F 12.3
16:00 258.9 F 19.7
18:00 270.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.7
1:00 272.5 F 28.3
2:00 271.9 F 28.1
3:00 271.9 F 27.8
4:00 272.7 F 28.5

Min 271.4 F 27.6
Max 272.7 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 218.1 F 0.0
10:00 230.2 F 0.0
15:00 227.6 F 0.0
20:00 220.4 F 0.0
25:00 219.8 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 18:33
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:57:05
CYCLE COMPLETE

11/08/2022 06:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____