
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002114
9:05 03/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:04
STERILIZING 3:08
136.5 C 223kPa 3:08
136.3 C 223kPa 3:38
136.3 C 224kPa 4:08
136.4 C 223kPa 4:38
136.3 C 223kPa 5:08
136.1 C 222kPa 5:38
136.4 C 224kPa 6:08
136.4 C 223kPa 6:38
136.4 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:28
CYCLE COMPLETE 37:28

Digital Signature #
89CF0BBDFBE294D3
