
Midmark M11 - v1.0.5
Total Cycles: 2035

Name: RITTER2
Sterilizer ID: V2313228
02/09/2023 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.7 F 0.0
2:00 163.2 F 0.1
4:00 186.5 F 3.4
6:00 208.1 F 8.0
8:00 223.3 F 13.5
10:00 231.4 F 9.7
12:00 235.0 F 9.0
14:00 242.0 F 11.5
16:00 246.9 F 14.1
18:00 259.9 F 21.4
20:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.6
2:00 271.5 F 29.0
3:00 273.0 F 29.8
4:00 272.1 F 29.2

Min 271.5 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 212.9 F 0.0
10:00 219.1 F 0.0
15:00 219.2 F 0.0
20:00 217.8 F 0.0
25:00 217.0 F 0.0
30:00 232.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

02/09/2023 06:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____