
Midmark M11 - v1.0.5
Total Cycles: 3297

Name: RITTER2
Sterilizer ID: V2313228
01/15/2024 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 78.7 F 0.0
4:00 99.1 F 0.0
6:00 105.2 F 0.7
8:00 143.1 F 3.4
10:00 173.4 F 6.6
12:00 212.2 F 10.5
14:00 210.6 F 5.7
16:00 222.8 F 6.9
18:00 232.2 F 10.7
20:00 231.3 F 7.7
22:00 239.5 F 11.3
24:00 236.2 F 9.0
26:00 249.1 F 15.1
28:00 260.3 F 21.3
30:00 270.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.7 F 29.3
2:00 271.3 F 28.8
3:00 272.4 F 29.4
4:00 271.4 F 28.6

Min 271.3 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 223.5 F 0.0
10:00 205.3 F 0.0
15:00 185.5 F 0.0
20:00 182.0 F 0.0
25:00 180.4 F 0.0
30:00 178.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

01/15/2024 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____