
Midmark M11 - v1.0.0
Total Cycles: 1815

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2022 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 165.5 F 3.2
4:00 186.8 F 7.3
6:00 206.6 F 12.4
8:00 222.6 F 6.3
10:00 235.5 F 11.4
12:00 233.6 F 8.1
14:00 241.7 F 11.3
16:00 245.1 F 12.7
18:00 257.7 F 19.3
20:00 268.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.3 F 28.2
3:00 272.6 F 28.9
4:00 271.6 F 28.2

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 228.6 F 0.0
10:00 224.6 F 0.0
15:00 223.2 F 0.0
20:00 204.1 F 0.0
25:00 217.4 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

04/12/2022 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____