
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001040
12:47 22/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:13
STERILIZING 4:28
136.4 C 223kPa 4:28
136.2 C 223kPa 4:58
136.1 C 219kPa 5:28
136.4 C 225kPa 5:58
136.3 C 223kPa 6:28
136.3 C 223kPa 6:58
136.3 C 223kPa 7:28
136.4 C 224kPa 7:58
136.4 C 224kPa 7:58
VENTING 7:58
AIR DRYING 8:52
CYCLE COMPLETE 8:54

Digital Signature #
461B3A368B291C48
