

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3748

Name: RITTER2  
Sterilizer ID: V2313228  
07/02/2024 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 70.7 F 0.0  
4:00 102.4 F 0.0  
6:00 110.2 F 0.2  
8:00 139.9 F 2.3  
10:00 166.2 F 4.8  
12:00 188.6 F 7.3  
14:00 200.9 F 10.9  
16:00 213.7 F 6.2  
18:00 224.3 F 8.0  
20:00 234.0 F 11.9  
22:00 231.6 F 8.0  
24:00 242.3 F 13.1  
26:00 240.6 F 11.1  
28:00 251.9 F 16.7  
30:00 262.0 F 22.6  
32:00 271.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.9 F 29.7  
2:00 271.4 F 28.9  
3:00 272.4 F 29.8  
4:00 272.1 F 29.3  
  
Min 271.2 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 194.8 F 0.0  
10:00 187.8 F 0.0  
15:00 183.0 F 0.0  
20:00 193.1 F 0.0  
25:00 197.8 F 0.0  
30:00 200.9 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:11:02  
CYCLE COMPLETE

07/02/2024 10:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_