
Midmark M11 - v1.0.0
Total Cycles: 1041

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.3 F 0.0
2:00 206.2 F 4.3
4:00 228.4 F 10.7
6:00 228.2 F 7.5
8:00 241.4 F 12.6
10:00 240.7 F 11.9
12:00 241.1 F 10.8
14:00 254.6 F 17.8
16:00 266.3 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.2
2:00 271.5 F 28.7
3:00 272.7 F 29.2
4:00 271.7 F 28.6

Min 271.5 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.3 F 0.0
5:00 229.7 F 0.0
Dry Cycle Aborted

FILLING: 2:58
HEATING: 17:33
STERILIZING: 4:00
VENTING: 3:14
DRYING: 9:37
TOTAL CYCLE: 00:37:22

07/21/2021 03:30 PM
ACCEPTED: X
REJECTED:
Operator: U01