

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001629  
10:46 29/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:40  
STERILIZING 3:52  
136.4 C 226kPa 3:52  
136.3 C 225kPa 4:22  
136.4 C 224kPa 4:52  
136.2 C 223kPa 5:22  
136.3 C 225kPa 5:52  
136.2 C 224kPa 6:22  
136.4 C 226kPa 6:52  
136.4 C 223kPa 7:22  
136.4 C 223kPa 7:22  
VENTING 7:22  
AIR DRYING 8:15  
CYCLE COMPLETE 8:27  
-----

Digital Signature #  
2A0812AD10D649A7  
-----