
Midmark M11 - v1.0.5
Total Cycles: 4436

Name: RITTER2
Sterilizer ID: V2313228
03/19/2025 03:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.4 F 0.0
2:00 115.0 F 1.0
4:00 132.5 F 3.2
6:00 166.7 F 7.1
8:00 186.8 F 10.9
10:00 206.8 F 4.2
12:00 218.7 F 8.2
14:00 231.1 F 12.7
16:00 232.9 F 9.2
18:00 239.4 F 10.4
20:00 244.5 F 13.4
22:00 256.4 F 19.7
24:00 267.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.5
1:00 272.1 F 30.0
2:00 272.6 F 29.9
3:00 271.5 F 29.4
4:00 272.5 F 30.4

Min 271.4 F 29.3
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 209.1 F 0.0
10:00 217.6 F 0.0
15:00 218.2 F 0.0
20:00 215.9 F 0.0
25:00 213.8 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:20
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:57
CYCLE COMPLETE

03/19/2025 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____