
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001142
12:49 23/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:15
STERILIZING 4:22
136.0 C 220kPa 4:22
136.0 C 222kPa 4:53
136.3 C 224kPa 5:23
136.4 C 221kPa 5:53
136.3 C 225kPa 6:23
136.3 C 224kPa 6:53
136.4 C 226kPa 7:23
136.4 C 225kPa 7:53
136.4 C 225kPa 7:53
VENTING 7:53
AIR DRYING 8:48
CYCLE COMPLETE 23:48

Digital Signature #
856FB6421016E3BC
