
Midmark M11 - v1.0.5
Total Cycles: 4114

Name: RITTER2
Sterilizer ID: V2313228
11/13/2024 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.5 F 0.0
2:00 148.7 F 0.2
4:00 158.9 F 3.0
6:00 187.6 F 6.5
8:00 212.5 F 10.6
10:00 221.3 F 5.8
12:00 229.0 F 9.5
14:00 235.2 F 10.4
16:00 241.5 F 11.8
18:00 240.0 F 10.6
20:00 252.6 F 16.9
22:00 262.9 F 23.1
24:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.2 F 29.4
2:00 271.6 F 29.4
3:00 272.5 F 29.6
4:00 272.5 F 30.0

Min 271.0 F 28.8
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 211.1 F 0.0
10:00 204.3 F 0.0
15:00 196.5 F 0.0
20:00 192.7 F 0.0
25:00 188.3 F 0.0
30:00 188.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

11/13/2024 03:04 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____