
Midmark M11 - v1.0.0
Total Cycles: 1844

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 73.7 F 0.0
4:00 91.9 F 0.0
6:00 131.9 F 0.0
8:00 151.0 F 0.4
10:00 176.4 F 3.7
12:00 207.6 F 7.0
14:00 231.0 F 10.2
16:00 226.7 F 10.3
18:00 227.8 F 7.1
20:00 237.1 F 11.0
22:00 233.6 F 7.4
24:00 243.8 F 12.5
26:00 240.7 F 10.4
28:00 252.9 F 16.4
30:00 263.4 F 22.6
32:00 272.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.6
2:00 271.2 F 28.1
3:00 272.5 F 28.7
4:00 271.4 F 28.0

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 232.4 F 0.0
10:00 210.9 F 0.0
15:00 216.7 F 0.0
20:00 225.7 F 0.0
25:00 231.3 F 0.0
30:00 229.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:16
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:10:20
CYCLE COMPLETE

04/20/2022 08:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____