
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000228
10:34 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:54
STERILIZING 5:29
136.0 C 226kPa 5:29
136.3 C 232kPa 5:59
136.4 C 233kPa 6:29
136.4 C 233kPa 6:59
136.3 C 232kPa 7:29
136.4 C 232kPa 7:59
136.1 C 232kPa 8:29
136.4 C 233kPa 8:59
136.4 C 233kPa 8:59
VENTING 8:59
AIR DRYING 10:00
CYCLE COMPLETE 18:51

Digital Signature #
CA3509BCC6CB1276
