
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005883
8:59 04/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:17
STERILIZING 4:28
136.4 C 224kPa 4:28
136.4 C 222kPa 4:58
136.4 C 222kPa 5:28
136.2 C 223kPa 5:58
136.2 C 223kPa 6:28
136.1 C 223kPa 6:58
136.1 C 218kPa 7:28
136.4 C 224kPa 7:58
136.4 C 224kPa 7:58
VENTING 7:58
AIR DRYING 8:51
CYCLE COMPLETE 9:20

Digital Signature #
CF8FE5A95485525A
