
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000326
11:12 11/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:39
STERILIZING 4:44
136.3 C 232kPa 4:44
136.4 C 233kPa 5:14
136.4 C 233kPa 5:44
136.3 C 233kPa 6:14
136.2 C 231kPa 6:44
136.3 C 231kPa 7:14
136.4 C 232kPa 7:44
136.4 C 233kPa 8:14
136.4 C 233kPa 8:14
VENTING 8:14
CYCLE ABORTED 8:48
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:48

Digital Signature #
77AA4D70F5D7F0D9
