
Midmark M11 - v1.0.5
Total Cycles: 91

Name: RITTER2
Sterilizer ID: V2313228
06/10/2021 03:39 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.4 F 0.0
2:00 157.8 F 2.6
4:00 165.7 F 6.3
6:00 189.0 F 10.9
8:00 213.9 F 5.3
10:00 232.0 F 10.7
12:00 231.0 F 8.1
14:00 241.9 F 11.6
16:00 245.9 F 13.5
18:00 258.6 F 20.2
20:00 269.7 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.4
1:00 272.8 F 29.4
2:00 272.7 F 29.6
3:00 272.2 F 30.0

Min 271.8 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 210.7 F 0.0
10:00 221.6 F 0.0
15:00 223.2 F 0.0
20:00 221.7 F 0.0
25:00 221.4 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:42
STERILIZING: 3:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

06/10/2021 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____